

File
LR1409

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Casella Holdings Limited	
Address	Regent House, Wolseley Road Kempston, Bedford Bedfordshire MK42 7JY United Kingdom	
Project Number	LR1409-3R3	
Product	Personal Air Sampling Pump	
Model Number	VAPex and VAPex Pro	
Electrical Ratings	Battery operated using battery pack P/N HCC802050-1S2P-L01, rated 3.7Vdc, 1600mAh, 5.92Wh	
Markings	Class I, Zone 0, AEx ia IIC T4 Ga Ex ia IIC T4 Ga Ta=-20°C to +45°C	
Applicable Standards	CSA C22.2 No. 61010-1, Ed. 3 CSA C22.2 No. 60079-0-19 CSA C22.2 No. 60079-11-14	UL 61010-1, Ed. 3 UL 60079-0:19 UL 60079-11:2013
Factory/Manufacturing Location	Casella Holdings Limited Regent House, Wolseley Road, Kempston, Bedford Bedfordshire MK42 7JY, United Kingdom	
Conditions of Certification	None	

Statement of Compliance: The product(s)/equipment identified in this Certificate and described in the Certification Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service



Issued By: Rob Kohuch, Certification Manager

Signature: *Rob Kohuch*

Date: April 21, 2026